Enbio S

Process times

All process times are drying included. Start-to-end, from instrument insertion to having them ready. Unwrapped FAST / 134°C

Wrapped Class B / 134°C Wrapped Class B / 121°C


7 min 15 min

30 min










Usable load space



Data archiving

All process data is automatically written to a USB disk.



Water supply

Enbio uses distilled water from an external tank.



Process reports

Dedicated software for process data reading and printing.



Power supply

The Enbio S works with any typical electrical outlet.





Class B

Vacuum Pouches+ drying



Medical Certified

Medical device class II b EN-13060 / 98/42/EEC Directive FDA Approved



Power supply	230V/50Hz
Installed power	2250 W
Max. power consumption	10 A

Pressure and temperature

Working pressure	2.1 Bar
Max. pressure	2.3 Bar
Max. process temperature	137°C

Miscellanous

Ingress protection	IP20
Data archiving	USB stick
Water conductivity*	< 15 µS/cm







www.icona.net.au



Distributed by:

Keilor Park Vic 3042



Switzerland